



Important safety information
Bridge Occlusion Balloon

Indications for Use

The Bridge Occlusion Balloon catheter is indicated for temporary vessel occlusion of the superior vena cava in applications including perioperative occlusion and emergency control of hemorrhage. Any use for procedures other than those indicated in these instructions is not recommended.

Contraindications

No known contraindications.

Potential Adverse Events

The adverse events associated with an occlusion balloon procedure include, but are not limited to the following: allergic reactions, death, embolization, hematoma, hemorrhage, sepsis/infection, short-term hemodynamic deterioration, thromboembolic episodes, vascular thrombosis, vessel dissection, vessel perforation, vessel spasm.

Caution: Federal law restricts this device to sale by or on the order of a physician.

