


IHE Integration Statement

Advanced Visualization Workspace V16



- This page is left intentionally empty -

1. Contents

1. Contents 3

2. Introduction..... 4

 2.1. Overview4

 2.2. Important Note to the Reader4

 2.3. General Acronyms and Abbreviations5

3. IHE Integration Statement..... 6

2. Introduction

2.1. Overview

The Integrating the Healthcare Enterprise (IHE) Integration Statement specifies the Integration Profiles, its Actors, and Options Philips Healthcare has chosen for implementing in this product. This document helps the reader to investigate whether and to what extent interoperability with other products might be supported.

The IHE Technical Framework identifies a subset of the functional components of the healthcare enterprise and specifies their interactions in terms of a set of coordinated transactions. The actors and transactions described in the IHE Technical Framework are abstractions of the real-world healthcare information system environment. While some of the transactions are traditionally performed by specific product categories (e.g. HIS, RIS, PACS, or modalities), the IHE Technical Framework intentionally avoids associating functions or actors with such product categories. For each actor, the IHE Technical Framework defines only those functions associated with integrating information systems. The IHE definition of an actor should therefore not be taken as the complete definition of any product that might implement it, nor should the framework itself be taken as the complete definition of a healthcare information system architecture.

This IHE Integration Statement provides the reader with a high-level view of supported IHE integration profiles. For further investigations, additional information can be found in the DICOM Conformance Statement of this product and in the IHE Technical Framework.

2.2. Important Note to the Reader

This IHE Integration Statement by itself does not guarantee successful interoperability of this Philips product with other products. The user (or user's agent) should be aware of the following issues:

Interoperability

Interoperability refers to the ability of application functions, distributed over two or more systems, to work successfully together. It is the user's responsibility to analyze thoroughly the application requirements and to specify a solution that integrates Philips equipment with non-Philips equipment.

Verification

Philips equipment has been carefully tested to ensure that the actual implementation of the IHE Integration Profiles corresponds with this Integration Statement.

Where Philips equipment is linked to other equipment, the first step is to compare the relevant Integration Statement. If the Integration Statement indicates that successful information exchange should be possible, additional verification tests will be necessary to ensure the functionality, performance, accuracy and stability of image and image related data. It is the responsibility of the user (or user's agent) to specify the appropriate test suite and to carry out the additional verification tests.

New versions of the IHE Technical Framework

The IHE Technical Framework will evolve in future to meet the user's growing requirements and to incorporate new features and technologies. Philips is actively involved in this evolution and plans to adapt its equipment to future versions of the IHE Technical Framework. In order to do so, Philips reserves the right to make changes to its products or to discontinue its delivery.

The user should ensure that any non-Philips provider linking to Philips equipment also adapts to future versions of the IHE Technical Framework. If not, the incorporation of IHE enhancements into Philips equipment may lead to loss of connectivity (in case of networking).

2.3. General Acronyms and Abbreviations

The following acronyms and abbreviations are used in the document.

- Actor An entity within a use case that performs an action
- DICOM Digital Imaging and Communication in Medicine
- HIS Hospital Information System
- HL7 Health Level 7
- IHE Integrating the Healthcare Enterprise
- PACS Picture Archiving and Communication System
- RIS Radiology Information System

3. IHE Integration Statement

Version history:

Version:	Date of issue:	Description of change
1.0	01-Dec-2025	Initial release of Advanced Visualization Workspace V16
2.0	11-Jun-2026	Updated for release candidate

Vendor	Product Name	Version
Philips Medical Systems	Advanced Visualization Workspace	V16
This product implements all the transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors and Options listed below.		

Integration Profiles Implemented	Actors Implemented	Options Implemented
Access to Radiology Information (ARI)	Image Display ¹	Multiple Sources
Scheduled Workflow (SWF)	Image Display ¹	None
Consistent Presentation of images (CPI) ²	Image Display ³	None
	Evidence Creator	None
Key Image Note (KIN)	Image Display ⁴	None
	Evidence Creator	None
NM Image (NMI)	Evidence Creator	Result screen export
	Image Display	General NM
		Cardiac NM
Portable Data for Imaging (PDI)	Image Display	MPR
	Portable Media Creator	DVD Media USB Media
	Image Display	DVD Media USB Media
Consistent Time (CT)	Print Composer	DVD Media USB Media
	Time Client	None
Audit Trail and Node Authentication (ATNA)	Secure Node ⁵	STX: No Secure Transport ATX: TLS Syslog ATX: UDP Syslog Radiology Audit Trail

- Advanced Visualization Workspace V16 complies to the IHE Profile, series retrieval is fully

- automated without the need of user interaction.
2. CPI Profile only applicable for Multi-Modality Viewer on Advanced Visualization Workspace V16.
 3. Signed Images may appear with suboptimal windowing when viewed on the Multi-Modality viewer on the Advanced Visualization Workspace V16. Users can use the “Automatic” Window preset from the dropdown options instead of the “Default” to achieve proper windowing. This option is accessible via the application tool panel.
 4. Advanced Visualization Workspace V16 complies with all the requirements of the IHE Profile, except supporting the following:
 - In study level query, Study Time (0008,0030) is not supported.
 - There is no user option to perform series and image level query.
 5. Advanced Visualization Workspace V16 complies to the IHE Profile, except the identification of multiple active participants for the DICOM Instance Transferred and Study Deleted actions.

IHE Conformance Statements	Philips IHE (www.philips.com/IHE)
DICOM Conformance Statements	Philips DICOM (www.philips.com/DICOM)
More about products from Philips Healthcare	https://www.philips.com/healthcare/about/customer-support
General information on IHE	http://www.ihe.net/ (International and North America) https://www.ihe-europe.net/ (Europe) http://www.ihe-j.org/en/ (Japan)

- This part of the page is left intentionally empty -

Issued by:

Philips Medical Systems Nederland B.V.

Veenpluis 6

5684 PC Best

The Netherlands

Note that the address is also known as “Veenpluis4-6”, which refers to the same physical location.

Internet: <https://www.philips.com/healthcare/about/customer-support>

Doc Id: 1967565

