

Where do we go from here? Insights from the 2018 IPEM Dose survey and opportunities for quality improvement 20th November-2019

Carl Horsfield

Principal Physicist-Radiotherapy Imaging; HCPC no: CS18532

phone: 01482 461371

email: carl.horsfield@hey.nhs.uk | website: hullrad.org.uk

Radiation Physics Department | The Queen's Centre | Castle Hill Hospital | Castle Road | Cottingham | HU16 5JQ

Acknowledgements

- Thanks go to:
 - Tim wood for providing slides from the “First UK survey of doses from radiotherapy CT scans”
 - Annie Tonks for her work on an internal RT CT dose audit.
 - Gill Millen for inviting me to give this talk.

Castle Hill Hospital Radiotherapy Team




<http://www.hullrad.org.uk/>



RADIOTHERAPY EQUIPMENT

- Six Varian Linacs, two Clinac iX systems, three Truebeam systems and a state of the art Halcyon system, all with portal MV imaging and kV CBCT imaging. Varian Real-time Position Management (RPM) system.
- Nucletron Flexitron-HDR unit
- Xstrahl 200 Superficial Therapy X-Ray (SXR) Unit
- Ariane Papillon 50 Contact Therapy X-Ray Unit
- 2 Philips Brilliance Big Bore 16-slice CT Scanners
- Varian Eclipse treatment planning system for external beam therapy and Oncentra Masterplan for brachytherapy
- IBA Compass pre treatment verification system

Our Department




National first for MPACE accreditation

Already compliant with **ISO9001** accredited by BSI.


Now accredited for **BS70000** which is a new standard specific to Medical Physics and Clinical Engineering services.



**Accredited to
BS 70000:2017**



Where we are coming from





IPEM

Institute of Physics and
Engineering in Medicine

The first UK survey of doses from radiotherapy treatment planning CT scans for adult patients

Tim Wood, Anne Davis, James Earley, Sue Edyvean,
Una Findlay, Rebecca Lindsay, Rosy Plaistow, Andrew
Nisbet, Antony Palmer, Matt Williams

Working party of Radiotherapy & Diagnostic
Radiology Special Interest Groups

The aims of the working party

- To fill in the 'gap' by facilitating sharing and comparison of dose information for radiotherapy imaging
- Or specifically...

The aims of the working party

- To undertake an audit of typical imaging doses for a range of common X-ray imaging procedures undertaken in Radiotherapy departments
 - This includes planning CT scans, on treatment CBCT imaging and tomotherapy (may also consider planar X-ray and fluoroscopy)
- To publish a range of typical 'doses' for common procedures
 - Like PHE do with national reference doses in diagnostic imaging
 - If data is good enough, this should allow adoption as national reference doses for RT imaging
- Make data available to the UK Radiotherapy community that will enable better optimisation of imaging
 - Including direct feedback to participating centres
- Identify best practice that will ultimately benefit patients

The 'grand plan'


- Aim for at least two peer-reviewed publications in PMB
 - **Planning CT doses**
(out now: <http://iopscience.iop.org/article/10.1088/1361-6560/aacc87/meta>)
 - **CBCT doses/tomotherapy**
(data collection soon, aiming for publication in 2019)
- Want to include a simple evaluation of **image quality** to go alongside the assessment of doses
 - Later, follow-up work if there is time for the working party

CT planning scans

- Launched data collection in February 2017, and final dataset came in September 2017
- For each patient:
 - CT Dose Index ($CTDI_{vol}$)
 - Dose-length product (DLP)
 - Scan length
 - Patient weight (if available)

Pre-data collection questionnaire

- With regards to **CT planning scan equipment and protocols**;
 - In the 59 UK centres to submit data, there are a total of 92 CT scanners currently 'in use'
 - Fairly evenly split between four vendors
 - The seven proposed 'clinical protocols' were used by the majority of centres
 - There was no common 'other' protocol suggested for audit
 - *Note*, no nodes was selected for audit due to more standardised protocols (scan lengths) being used




CT Protocol	N	% of centres
Breast (no nodes)	57	96.6
Prostate (no nodes)	56	94.9
Gynae (no nodes)	46	78.0
Lung 3D	53	89.8
Lung 4D	48	81.4
Brain	44	74.6
Head and neck	54	91.5


Data analysis

- For this study, **median** from each scanner was used to define scanner average $CTDI_{vol}$, DLP, scan length (not mean)
 - More robust against outliers e.g. very obese patients
 - In accordance with draft guidance from the ICRP on '*Diagnostic Reference Levels in Medical Imaging*' and is also a widely used technique in many centres
 - For data plots, error bars are the 95% confidence intervals
- **Third quartile** ('national reference') and **median** ('achievable') of the **scanner median data** were calculated in Excel


Breast – DLP




Breast – FBP vs iterative



Breast – Scan length




Breast – AEC vs fixed mAs



Lung 4D

- Relatively small data set
 - 41 DLP scanner medians
 - 40 CTDI_{vol} scanner medians (39 excluding max CTDI systems)
 - 36 scan length medians
- Very wide range of doses
 - Factor of 18.6 between min and max DLP
 - Factor of 16.7 between min and max CTDI_{vol}
- Very strong vendor dependence
 - Different approaches to 4D CT data acquisition and processing?



* Indicates maximum CTDI on older Toshiba scanners using AEC

Proposed reference levels

CT Planning scans – Reference Values

(3rd quartile)

Examination	Phantom diameter (cm)	CTDI _{vol} (mGy)	DLP (mGy.cm)	Scan length (mm)
Breast	32	10	390	360
Gynaecological	32	16	610	400
Lung 3D	32	14	550	390
Lung 4D	32	63	1750	340
Prostate	32	16	570	340
Brain	16	50	1500	290
Head and neck	16	49	2150	420

CT Planning scans – Achievable values

(Median)

Examination	Phantom diameter (cm)	CTDI _{vol} (mGy)	DLP (mGy.cm)	Scan length (mm)
Breast	32	8	280	330
Gynaecological	32	12	510	390
Lung 3D	32	10	410	370
Lung 4D	32	36	1170	330
Prostate	32	13	420	310
Brain	16	42	1110	250
Head and neck	16	26	1080	400

Standard Practice

- Local annual dose audit

Local annual dose audit

- Patient data was collected automatically via the OpenREM dose management system from an approximate six month period (end of July 2018 to January 2019), detailing the $CTDI_{vol}$, DLP and scan length for each patient examination across two radiotherapy treatment planning CT scanners, CT1 and CT2, at CHH.
- The data was collected and reviewed in Microsoft Excel and categorised into the seven examination protocols of breast, gynaecological, lung 3D, lung 4D, prostate, brain and head and neck, as per the NDRLs and with an additional examination protocol for DIBH breast. The median values for $CTDI_{vol}$, DLP and scan length were then calculated alongside deviations from the NDRL values for each of the two Philips machines.
- Additional achievable levels were proposed in the IPEM report based on the median of the scanner median distributions of the 68 radiotherapy CT scanners included in the audit




Figure 1: Patient dose audit data for the $CTDI_{vol}$ for the eight examination protocols. Each protocol includes a measure for each of the two machines, CT1 and CT2, with the relevant NDRL and achievable level for comparison. Note: the brain and head and neck examination protocol values for $CTDI_{vol}$ and DLP have been scaled to match the 16 cm phantom used for the NDRLs.

Credit to Annie Tonks for conducting this audit




Figure 2: Patient dose audit data for the DLP for the eight examination protocols. Each protocol includes a measure for each of the two machines, CT1 and CT2, with the relevant NDRL and achievable level for comparison. Note: the brain and head and neck examination protocol values for $CTDI_{vol}$ and DLP have been scaled to match the 16 cm phantom used for the NDRLs.

Credit to Annie Tonks for conducting this audit




Figure 3: Patient dose audit data for the scan length measurements for the eight examination protocols. Each protocol includes a measure for each of the two machines, CT1 and CT2, with the relevant NDRL and achievable level for comparison. Note: the brain and head and neck examination protocol values for $CTDI_{vol}$ and DLP have been scaled to match the 16 cm phantom used for the NDRLs.

Credit to Annie Tonks for conducting this audit


Table 1: Local DRL values based on the 32 cm diameter phantom (size used for all protocols on the scanners).

Examination	CTDI _{vol} LDRL (mGy)	DLP LDRL (mGycm)	Scan Length LDRL (mm)
Breast	7	230	330
Gynaecological	10	410	390
Lung 3D	8	340	370
Lung 4D	24	800	320
Prostate	10	330	320
Brain	22	590	230
Head and Neck	8	320	390

Reception of Images

- What do our doctors actually think of image quality.
- What can we do to improve?

Reception of Images



Activity : 'Volume and Prescribe Urology IMRT' for Patient 'Prostate, Nonclincial1 (RAProstate1)' - Activity Status : 'Completed' - Proce...

Patient Status: Test Patient Patient I/O Status: Outpatient Prim. Oncologist: Course: Payor: Hospital: Castle Hill Hospital Department: RadiotherapyCHH Covering Provider:

Related Diagnoses

Related	Type	Onset Date	Code	Clinical Description	Sequence	Ran...
<input checked="" type="checkbox"/>	Primary	31/03/2019	C61	Malignant neoplasm of prostate	1	1 - ...
<input checked="" type="checkbox"/>	Primary	02/02/2017	Referral_1	Referral received date for Course 1	2	1 - ...

Procedure Code Details


	Qty	Procedure Code	Account Number	Exportable	Code Modifier	Completed by	Date of Service ...	Grou
<input type="checkbox"/>	1	HEY-TP-068 - High quality image		<input type="checkbox"/>		Carl Horsfield	12/11/2019 10:2...	Onc
<input type="checkbox"/>	1	HEY-TP-069 - Acceptable image		<input type="checkbox"/>		Carl Horsfield	12/11/2019 10:2...	Onc
<input type="checkbox"/>	1	HEY-TP-070 - Low quality image		<input type="checkbox"/>		Carl Horsfield	12/11/2019 10:2...	Onc
<input type="checkbox"/>	1	HEY-TP-071 - Rejected image		<input type="checkbox"/>		Carl Horsfield	12/11/2019 10:2...	Onc

Additional Activity Attributes


Image Quality Problem						

Help Previous Patient Next Patient Save Edits Cancel Edits

Reception of Images




Stated Reason for Low Quality




Reception of Images

- There is some work to be done.
- We could increase dose to help with Streaks, and noise. But Urology are already close to the Reference Level.



Where we going to?



Any Questions?

