PHILIPS
MEDICAL
SYSTEMS
Page 1 of 6
2009 July

Product name:		Gemini TF Big Bore PET Gantry		
Identification code(s)		4535-674-7985x (x) Denotes revision level wildcard designation. (Numbers range from 1-9)		
Total weight (in Kg) 90		907kg		
Producer/	Name company:		Philips Medical Systems (Cleveland), Inc.	
Manufacturer	Address:		595 Miner Road, Cleveland, Ohio	
	Zip code:		44143	
	Countr	y:	United States of America	

Recycle Info	Items:	Location (Qty.)
Special attention	None	
Fluids / Gases	None Items:	Location (Qty.)
Batteries To be Removed	Type: Sealed Lead Acid - PET Gantry UPS Lithium Ion Button Cell	Location (Qty.) See Figure 2 (1) See Figure 1 (1)
10 be Kellioved		
Hazardous To be Removed	Substances: Printed Circuit Board Assembly (PCBA) (*) – Power Supply Printed Circuit Board Assembly (PCBA) (*) – Power Dist Printed Circuit Board Assembly (PCBA) (*) Computer Fan & Motor Control Panels Lead (Pb) – Shielding Printed Circuit Board Assembly (PCBA) (*) Printed Circuit Board Assembly (PCBA) (*) - PMTs	Location (Qty.) See Figure 1 (1) See Figure 1 (7) See Figure 1 (7) See Figure 1 (1) See Figure 1 (1) See Figure 1 (2) See Figure 1 (1) See Figure 1 (1) See Figure 1 (1) See Figure 1 (420)
	Printed Circuit Board Assembly (PCBA) (*)	See Figure 2 (3)

PHILIPS	
MEDICAL	Page 2 of 6 2009 July
SYSTEMS	2009 July

	Fan & Motors	See Figure 2 (2)
	Printed Circuit Board Assembly (PCBA) (*)	See Figure 2 (1)
	Printed Circuit Board Assembly (PCBA) (*) - Power Supply	See Figure 2 (3)
	Printed Circuit Board Assembly (PCBA) (Pb) – (inside Cardcage)	See Figure 2 (12)
	Printed Circuit Board Assembly (PCBA) (*) – Motor Controller	See Figure 3 (3)
	* This gantry contains RoHS compliant and non-RoHS compliant PCBAs and components. Refer to the markings on the component for recycling purposes.	
Radioactive	Substances:	Location (Qty.)
	Naturally Occurring Radioactive Material (NORM) – Scintillation Crystal	See Figure 1 (28336 crystals)
To be Removed	This gantry incorporates Lu-176 LYSO crystal detectors. Lu-176 is primarily a beta-emitting isotope. These particles are completely absorbed by the crystals themselves. There are additionally some low-to-medium energy gamma rays that may	e e
To be Kemoveu	escape the crystals. Therefore, a dose rate survey was performed to detect and measure this gamma radiation around the unit. As measurements observed were indistinguishable from background and no exposure to personnel is expected from the LYSC	d !,

PHILIPS Page 3 of 6 **MEDICAL** 2009 July **SYSTEMS**




Fan & Motor

PCBA (7)

Dist (inside)

Power Supply

Computer











Shielding




PCBA



Figure 1, PET Gantry Front

Page 4 of 6 2009 July

SYSTEMS



PHILIPS
MEDICAL
SYSTEMS
Page 5 of 6
2009 July




Figure 3, PET Gantry Rear Close-up