



## Suggested protocol

### Philips Respironics CoughAssist E70

Please refer to the CoughAssist E70 user manual for complete product descriptions, including indications and contraindications for use. Once it has been determined that the CoughAssist treatment is clinically appropriate, the following may be used as a suggested protocol. Please review the entire protocol before initiating therapy. This protocol is not intended as a substitute for advice from a licensed physician or other healthcare professional, and the prescription issued by the patient's physician should be followed.

#### Indications

For use with patients unable to cough or clear secretions effectively due to reduced peak expiratory flow.

- Those who might benefit from the use of the CoughAssist E70 include patients with an ineffective cough due to muscular weakness or dystrophy, myasthenia gravis, poliomyelitis, or other neurologic disorder with some paralysis of the respiratory muscles, such as spinal cord injury. It may also be used to treat ineffective cough due to other bronchopulmonary diseases, such as emphysema, cystic fibrosis, and bronchiectasis. It is effective for both trach and noninvasive use.
- In pediatric patients who are able to perform a reproducible forced expiratory flow maneuver, a value less than 50% of predicted is an indication for mechanical insufflation-exsufflation. PCF values > 270 LPM<sup>1</sup> have been established as the minimum level necessary to clear secretions in adults, but values for small children have yet to be validated and may be lower than the values for adults.<sup>2</sup>

#### Contraindications

- Any patient with a history of bullous emphysema
- Susceptibility to pneumothorax or pneumomediastinum
- Recent barotraumas

The above contraindications should be carefully considered before use.

Patients known to have cardiac instability should be monitored for pulse and oxygen saturation very closely<sup>3</sup>.

#### Warnings and cautions

Refer to the CoughAssist E70 user manual.

#### Patient preparation

The CoughAssist E70 should be carefully introduced to the patient. If treating a pediatric patient, a parent or trusted caregiver should be involved, if possible. Allow the patient to become familiar with the mask, especially if they are unaccustomed to positive pressure therapy.

## Implementation of CoughAssist E70

- Attach the CoughAssist patient circuit to the CoughAssist output, including a bacterial/viral filter, smoothbore tubing, and an appropriate interface (mask, mouthpiece, or trach adapter). If a mask is used, it should be of appropriate size to provide a tight seal. When used with a trach, attaching directly to an inline suction catheter allows for easy removal of secretions from the top of the trach.
- Explain principles of the CoughAssist E70 to the patient and the caregiver—deep inflation of the lungs followed by a forced exhalation of air aimed at removing secretions located in the central airways. Explain that the patient should relax so the air delivered by the CoughAssist E70 slowly expands the lung and chest wall. After the deep breath is delivered, the exhalation phase will begin immediately. Instruct the patient to exhale fully during the exhalation phase.<sup>2</sup>

## Settings and modes

### Manual mode

This mode may be used for initial acclimation to the CoughAssist E70 and for titrating pressures and adjusting times prior to using the Automatic mode.

- Begin with inspiratory pressures between +10 and +15 cm H<sub>2</sub>O and expiratory pressures of between –10 and –15 cm H<sub>2</sub>O to allow an introduction/acclimation to the device.<sup>4</sup> Set device inhale flow to the low setting.
- Press the “Therapy” button to start treatment. Position the appropriate interface to the patient. Start with a single cough cycle to allow for acclimation to the device. A cough cycle is one inspiration, one expiration, and then a pause, if needed. Move the Manual switch to the inhale position and hold for 1 to 3 seconds. Immediately move the Manual switch to the exhale position and hold for 0.5 to 2 seconds, then release the switch to the neutral position. Verify patient comfort and tolerance of the maneuver. Adjust timing to coordinate with the patient’s breath rate and rhythm.<sup>5</sup>
- Continue with several cough cycles, 3 to 6 for pediatric patients and 4 to 6 for adults, in a session. If desired, a pause of 2 to 5 seconds between cough cycles may be used. Confirm continued tolerance of the maneuvers. A rest period of 30 to 60 seconds may be used before the session is repeated. Return the patient to his/her normal oxygen or ventilation settings during the rest period, if necessary. Perform at least 3 to 5 sessions for pediatric patients and 4 to 6 sessions for adults for each treatment.<sup>2,5</sup>
- Gradually increase the inspiratory and expiratory pressures. Adjust inhale flow, if needed, for patient comfort. Continually monitor the patient for comfort and tolerance. Positive pressure levels can be established by evaluation of chest wall expansion and auscultation for bilateral air entry.<sup>5</sup> The displayed tidal volumes may be used to titrate inspiratory pressure levels to achieve adequate inspired volumes. The displayed values for peak cough flows may also be used to titrate expiratory pressure levels and to coach patient effort. Inspiratory and expiratory pressures of up to +/- 40 cm H<sub>2</sub>O show the best results and are generally well tolerated.<sup>2</sup>

- Whenever possible, solicit feedback from the patient regarding pressures and inspiratory and expiratory times. For example, a gesture of up or down can be used to indicate whether to increase or decrease pressures or times.
- Subsequent treatment sessions can be initiated at previously established pressures, times and flow rates. These settings may also be entered as the presets for use with the Automatic mode.

## Automatic mode

Automatic mode provides a timing feature that will automatically trigger to inspiration and cycle to expiration instead of manually moving the switch. Inhale and exhale times entered into the device will replace manually moving the switch. Cough-Trak is a feature in the Automatic mode that will synchronize inhalation with patient effort.

- Select Automatic mode on the display. Evaluate inspiratory drive of the patient and select the Cough-Trak On/Off setting accordingly. If pressures, times, and flows were titrated using the Manual method, use those final values as starting values for the Automatic mode. If titrating in Automatic mode, use the same initial settings as explained previously in the Manual mode paragraph and adjust for patient comfort and tolerance. If Cough-Trak is enabled, the pause time is determined by the patient trigger.
- Position the patient interface to the patient and start therapy. Therapy will start automatically if Cough-Trak is turned off. If Cough-Trak is enabled, therapy will start as soon as the patient initiates a breath.
- Adjustments to therapy can be made while in “Standby” or “Therapy” mode.
- An Oscillation feature, available in manual and automatic mode, can be set either during one or both phases of the cough cycle (insufflation and/or exsufflation). The aim of oscillations is to enhance loosening and mobilization of secretions. Start at a high frequency (20 Hz) and low amplitude (1 cm H<sub>2</sub>O) and adjust the settings to patient comfort.

When using CAE70 noninvasively begin oscillations in the inhale phase and assess tolerance of the oscillations. If desired, oscillations may also be used in the exhale phase if tolerated.

## Treatment length and process

- A cough cycle is composed of one inspiration, expiration, and pause phase. A standard sequence consists of 3 to 6 consecutive cough cycles for pediatric patients and 4 to 6 consecutive cough cycles for adults, followed by a rest period of 30 seconds. Patients should be returned to their normal oxygen or ventilator settings during the rest period, if necessary. Sequences can be repeated 3 to 6 times if needed to clear secretions.<sup>2,5</sup>
- The CoughAssist E70 manoeuvre may be ended on a positive pressure to preserve lung volume<sup>6</sup>.

- Suction equipment should be available and visible secretions should be removed via suction from mouth, tracheostomy tube, or tubing, as needed.<sup>5</sup>
- A properly applied abdominal thrust or lower chest compression, coordinated with the exsufflation phase of the cough cycle, can enhance peak cough flows and secretion clearance.<sup>7</sup> In manual mode, a foot pedal accessory, that overrides the manual lever, can allow the clinician to free one hand. This way the same clinician can manually deliver the therapy while applying the abdominal thrust.

## Use with a tracheostomy

Higher exhale pressures may be required to overcome the increased resistance of a tracheostomy or endotracheal tube. If the tracheostomy tube is cuffed, it is advised to have the cuff inflated for the CoughAssist E70 treatment and to use a means for trapping any secretions that may potentially accumulate in the treatment circuit. Standard water traps, sputum traps, or extension tubing with corrugated inner walls can serve well for this purpose.

NOTE: The FDA defines pediatric in subgroups as follows: newborn—from birth to 1 month of age; infant—greater than 1 month to 2 years of age; child—greater than 2 to 12 years of age; adolescent—greater than 12 to 21 years of age.

British Thoracic Society guidelines defines young children as children under 10 years of age.

<sup>1</sup> Bach JR, Ishikawa Y, Kim H. Prevention of pulmonary morbidity for patients with Duchenne Muscular Dystrophy. *Chest* 1997;112:1024-1028.

<sup>2</sup> [Fauroux B. et al. Physiological benefits of mechanical insufflation-exsufflation in children with neuromuscular diseases. \*Chest\* 2008;133:161-168.](#)

<sup>3</sup> [Bach JR. Update and perspective on noninvasive respiratory muscle aids. Part 2: the expiratory aids. \*Chest\* 1994;105:1538-44.](#)

<sup>4</sup> CoughAssist 3000-3200 User Manual and CoughAssist E70 Quick Start Guide

<sup>5</sup> [Miske, Laura J., et al. Use of the mechanical in-exsufflator in pediatric patients with neuromuscular disease and impaired cough. \*Chest\* 2004;125:1406-1412.](#)

<sup>6</sup> [Jeremy Hull, Roona Aniapravan, Elaine Chan, et al. British Thoracic Society guideline for respiratory management of children with neuromuscular weakness. \*Thorax\* June 2012 67: i1-i40.](#)

<sup>7</sup> [Bach JR. Mechanical insufflation-exsufflation comparison of peak expiratory flow with manually assisted and unassisted coughing techniques. \*Chest\* 1993;104:1553-1562.](#)

